

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis

BS 1016

Certified Reference Material for Carbon Steel Grade 1016 - UNS Number G10116

Estimated Analysis ¹		Estimated Analysis ¹	
Analysis listed as percent by weight			
Al	0.02	Nb	0.001
As	0.006	Ni	0.11
B	<0.005	O	<0.05
C	0.17	P	0.01
Ca	<0.005	Pb	<0.05
Co	0.02	S	0.03
Cr	0.09	Si	0.19
Cu	0.15	Sn	0.01
Fe	[98.3]	Ti	0.001
Mg	<0.005	V	0.001
Mn	0.77	W	<0.05
Mo	0.04	Zr	<0.005
N	<0.05		

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a hexagon disc approximately 60 mm and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by January 12, 2019.