

Brammer Standard Company, Inc.

Revised*

Certificate of Analysis

BS 17500

**Certified Reference Material¹ for Grade 17500 Copper Alloy
(UNS Number C17500)**

Certified Value²	Estimate of Uncertainty³		Certified Value²	Estimate of Ucertainty³
Analysis listed as percent by weight				
Al	0.0210	0.0007	Sb	0.0001 0.00004
Be	0.43	0.015	Si	0.0641 0.0025
C	0.0020	0.0003	Zn	0.0065 0.0010
Ca	0.0054	0.0005		
Co	2.31	0.05	Information Values⁴ - not certified	
Cr	0.0015	0.0002	<i>Ag</i>	<i>(0.0012)</i>
Fe	0.0262	0.0012	<i>As</i>	<i>(0.0002)</i>
Mg	0.0076	0.0012	<i>Cu</i>	<i>(97.1)</i>
Mn	0.081	0.002	<i>O</i>	<i>(0.0005)</i>
Ni	0.095	0.003	<i>S</i>	<i>(0.0001)</i>
P	0.0031	0.0006	<i>Sn</i>	<i>(0.0002)</i>
Pb	0.0005	0.00015	<i>Te</i>	<i>(0.0004)</i>

¹ Brammer Standard Company, Inc., is accredited to ISO Guide 34 as a Reference Material Producer to produce Certified Reference Materials by A2LA (Certificate Number 656.02)

² The certified value listed is the present best estimate of the true value based on the results of an interlaboratory testing program.

³ The uncertainties listed are based on value judgments of the material inhomogeneity and the 95% confidence interval. The half-width confidence interval C(95%) is shown on page 2.

⁴ The values listed in parentheses are not certified and are for information only.

* REVISION: Additional laboratory data was used to refine the certified values for Be, Co, Ni, Si, and Zn.

See the following pages for more information.

Certificate Number REV17500-071608p1

**Brammer Standard Company, Inc., 14603 Benfer Road, Houston, TX 77069-2895
Telephone (281) 440-9396 Fax (281) 440-4432 brammerstandard.com**

Analysis	* Al	* Be %	* C	* Ca	* Co %	* Cr	Fe	* Mg	* Mn %	* Ni %	* P
1	4 204	3 0.4128	1 18	4 48.4	3 2.25	3 13	3 250	7 58	4 0.0798	7 0.0925	10 26
2	7 208	3 0.42	1 19	3 53.8	3 2.255	3 14	3 256	7 68	3 0.0802	3 0.0928	10 27
3	3 209	4 0.427	1 20	3 54	3 2.322	7 14	3 257	9 70	3 0.0803	3 0.094	3 27
4	3 209	4 0.433	1 20	3 54.1	3 2.33	3 14.7	3 261	3 83	3 0.0809	3 0.0943	3 27
5	3 214	3 0.4340	1 21.6	3 55	4 2.340	3 15.0	3 271	3 83.5	3 0.0824	7 0.0977	3 30
6	3 215	3 0.4390	1 22.7	3 58	3 2.346	9 18	4 274	3 84	3 0.0845	3 0.0980	7 31
7		3 0.4400			3 2.36			3 84.4		3 0.0981	3 38
8											4 39
Average	209.8	0.4294	20.2	53.9	2.3147	14.8	261.5	75.8	0.08135	0.09534	30.6
Std Dev	4.1	0.0101	1.7	3.1	0.044	1.7	9.3	10.5	0.00179	0.00251	5.2
Certified	210	0.43 %	20	54	2.31 %	15	262	76	0.081 %	0.095 %	31
t	2.5706	2.4469	2.5706	2.5706	2.3646	2.5706	2.5706	2.4469	2.5706	2.4469	2.3646
C(95%)	4.3	0.0093	1.8	3.3	0.041	1.8	9.7	9.7	0.00188	0.00232	4.3

* indicates method of analysis

Analysis	* Pb	* Sb	* Si	* Zn	* Ag	* As	* Cu %	* O	* S	* Sn	* Te
1	7 4	7 0.80	3 610	3 59	7 7.2	6 1.0	12 96.9	2 2	9 0.51	7 1	8 1
2	7 4.3	8 1	3 622	3 59	3 9	6 1.1	5 96.9745	2 3	1 1	9 1.1	7 1.3
3	7 4.5	7 1.1	3 624	7 59	7 9.1	9 1.1	11 97.19	2 5.4	4 1	7 1.2	7 5.2
4	7 4.6	7 1.2	3 642	3 60	7 9.1	7 2.0	11 97.21	2 5.7	1 2.0	7 1.7	7 6.8
5	8 5	7 1.2	3 654	4 62	9 11	8 3		2 9.0	1 2.2	7 2.1	
6	7 5.1	9 1.5	3 663	7 62	3 15	3 7				3 4	
7	3 6		7 670	6 71.9	7 20					3 5	
8	4 7			6 76.1							
9				9 77							
Average	5.1	1.1	640.7	65.1	11.5	2.5	97.07	5.0	1.3	2.3	3.6
Std Dev	1.0	0.2	22.8	7.6	4.5	2.3	0.15	2.7	0.7	1.6	2.9
Certified	5	1	641	65	(12)	(2)	(97.1 %)	(5)	(1)	(2)	(4)
t	2.3646	2.5706	2.4469	2.306	2.4469	2.5706	3.1824	2.7764	2.7764	2.4469	3.1824
C(95%)	0.8	0.2	21.0	5.9	4.2	2.4	0.25	3.4	0.9	1.5	4.6

$C(95\%) = (t \times sd) / \sqrt{n}$ The half-width confidence interval, where t is the appropriate Student's t value, sd is the interlaboratory standard deviation, and n is the number of acceptable mean values. For further information regarding the confidence interval for the certified value see ISO Guide 35:1989 section 4.

* Methods of Analysis

Code	Method	Code	Method
1	Combustion - Inferred Absorption Method	7	AES-ICP-MS Inductively Coupled Plasma with Mass Spectrometry
2	Fusion - Thermal Conductivity Method	8	Graphite Furnace - Atomic Absorption Spectrometry
3	AES-ICP Inductively Coupled Plasma Spectrometry	9	Colorimetric method
4	AES- Spark Atomic Emission Spectrometry	10	Spectrophotometric method
5	Gravimetric method	11	X-Ray Fluorescence spectrometry
6	Flame Atomic Absorption Spectrometry	12	Titration method

AES = Atomic Emission Spectrometry

Co-operating Laboratories: The co-operating laboratories were:

Laboratory

Andrew S. McCreath & Son, Inc., Harrisburg, Pennsylvania
 Brammer Standard Co., Inc., Houston, Texas
 China National Analysis Center for Iron and Steel, Beijing, China
 Freeport-McMoran Copper & Gold, El Paso, Texas
 J. Dirats and Co., Inc., Westfield, Massachusetts
 Laboratory Testing Inc., Hatfield, Pennsylvania
 LECO Corporation, St. Joseph, Michigan
 NSL Analytical, Cleveland, Ohio
 Northern Analytical Laboratory, Inc., Merrimack, New Hampshire
 VHG Laboratories, Inc., Manchester, New Hampshire

Contact person

Robert E. Kozicki
 Richard P. Beaumont
 Prof. Wang Haizhou
 Jean Fernandez
 Eric E. Dirats
 Rick Heist
 Dennis Lawrenz
 Carm D'Agostino
 William A. Guidoboni & Peter S. Dickson
 Julie M. McIntosh

Certification Process: The requirements of ISO Guide 31, ISO Guide 34, ISO Guide 35, and ASTM Standard Guides E 1724 and E 1831 were followed for the preparation of this reference material and certificate of analysis. This is a Certified Reference Material as defined by ISO Guide 30.

Analysis: Chemical analyses were made on chips prepared by a lathe from the certified portion of the discs in accordance with ASTM Standard Practice E 255-07. The laboratories participating in the testing normally followed the requirements of ISO Standard 17025. Individual values listed on page 2 are the average of each analyst's results. Methods of analysis are listed on page 2.

Outliers: Some outlying data was excluded from the data listed on page 2 due to technical assessment of the cooperating laboratories and statistical evaluation.

Traceability: The following Certified Reference Materials were used to validate the analytical data listed on page 2: NIST SRMs 3101a, 3103a, 3105, 3109a, 3112a, 3113, 3128, 3131a, 3132a, 3136, 3139a, 3150, 3151, 3156, 3161a, 3168a;

Homogeneity: This Certified Reference Material was tested for homogeneity using ASTM Standard Method E 826 and found acceptable. It was also examined by optical emission spectrometry using and found to be compatible with the following Reference Materials: NIST SRM C1221, C1222, C1223; CTIF 4583, 4584, 4594, 4763, 4766, 4868, 4872, 4873; BS 172Be, 172Be-1.

Validity statement: ISO Guide 31 states that the certification should contain an expiration date for all materials where instability has been demonstrated or is considered possible, after which the certified value is no longer guaranteed by the certifying body. Whereas this material is in a solid form and stable, no expiration date is specified.

Source: This material was purchased from New Southern Resistance Welding, Inc., Pelham, Alabama..

Form: This Certified Reference Material is in the form of a disc, approximately 38 mm (1.50 inches) diameter and 12 mm (0.50 inches) thick.

Use: This Certified Reference Material is intended for use in optical emission and x-ray spectrometric methods of analysis. Refer to ISO Guide 33 for information about the use of Reference Materials.

Certified area: The entire depth of the disc may be used.

Caution: As with any bar material, avoid optical emission spectrometric burns in the center of the disc (5 mm radius), as some segregation may be present.

Sample Preparation: For best analytical results, use the same method for preparing the analytical surface on all reference materials as you use for production specimens. Avoid overheating the disc during surface preparation.

Certificate Number: The unique identification number for this certificate of analysis is REV17500-071608-px, where x indicates the page number. Refer to future Brammer Standard Company catalogs for information on any revisions to this or other Brammer Standard reference materials. You may also obtain information on revisions of certificates from the internet at www.brammerstandard.com.

Safety Notice: A Material Safety Data Sheet (MSDS) is not required for this material. This material will not release or otherwise result in exposure to a hazardous chemical under normal conditions of use. Inquiries concerning this Reference Material should be directed to:

Brammer Standard Co., Inc.	Phone: (281) 440-9396	web	www.brammerstandard.com
14603 Benfer Road			
Houston, Texas 77069-2895 USA	Fax: (281) 440-4432	e-mail	contact@brammerstandard.com

Certified by: _____ on July 16, 2008
Beau R. Brammer

Certificate Number REV17500-071608p3

**Brammer Standard Company, Inc., is accredited to ISO Guide 34 as a Reference Material Producer for the production of Certified Reference Materials and Reference Materials by A2LA (Certificate Number 656.02)
The scope of accreditation is listed on the website: www.brammerstandard.com**

By Certificate Number 10539, the Quality System of Brammer Standard Company, Inc., is registered to ISO 9001:2000 by National Quality Assurance, U.S.A.

**Brammer Standard Company's Chemical Laboratory is accredited to ISO Standard 17025 by A2LA.
(Certificate Number 656.01)**

References:

*ASTM documents available from ASTM, 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959,
Telephone: 610-832-9500 Fax: 610-832-9555 e-mail: service@astm.org Website: www.astm.org*

E 255-07 Standard Practice for Sampling Copper and Copper Alloys for the Determination of Chemical Composition

E 826 - 08 Standard Practice for Testing Homogeneity of a Metal Lot or Batch in Solid Form by Spark Atomic Emission Spectrometry

E 1724 - 95 (Reapproved 2001) Standard Guide for Testing and Certification of Metal and Metal-Related Reference Materials

E 1831 - 96 (Withdrawn 2005) Standard Guide for Preparing Certificates for Reference Materials Relating to Chemical Composition of Metals, Ores, and Related Materials.

ISO Guides and Standards available from Global Engineering - www.global.ihs.com

ISO Standard 17025 (Second edition, 2005), General requirements for the competence of calibration and testing laboratories.

ISO Guide 30 (Second edition, 1992), Terms and definitions used in connection with reference materials.

ISO Guide 31 (Second edition, 2000), Reference materials -Contents of certificates and labels.

ISO Guide 33 (Second edition, 2000), Uses of certified reference materials.

ISO Guide 34 (Second edition, 2000), General requirements for the competence of reference material producers.

ISO Guide 35 (Third edition, 2006), Certification of reference materials - General and statistical principles.

Other useful documents available from NIST, U.S. Department of Commerce, Gaithersburg, MD 20899.

NIST Special Publication 260-100, Handbook for SRM Users

NIST Special Publication 829, Use of NIST Standard Reference Materials for Decisions on Performance of Analytical Chemical Methods and Laboratories

Certificate Number REV17500-071608p4