

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis BS 1762

Certified Reference Material for Low Alloy Steel

Estimated Analysis ¹		Estimated Analysis ¹	
Analysis listed as percent by weight			
Al	0.05	Ni	1.16
As	0.02	O	<0.05
B	0.005	P	0.032
C	0.36	Pb	0.009
Ca	<0.005	S	0.036
Co	0.065	Sb	0.02
Cr	0.93	Si	0.37
Cu	0.14	Sn	0.079
Fe	[93.9]	Ta	0.06
Mg	<0.005	Ti	0.096
Mn	2.05	V	0.19
Mo	0.35	W	0.031
N	0.017	Zn	<0.05
Nb	0.074	Zr	0.014

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 37 mm in diameter and 25 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by June 21, 2021.