

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis

BS 254

Certified Reference Material for Alloy 254SMO - UNS Number S31254

Estimated Analysis ¹		Estimated Analysis ¹	
Analysis listed as percent by weight			
Al	0.008	Ni	18.4
As	<0.05	O	<0.05
B	0.002	P	0.025
C	0.019	Pb	<0.005
Ca	<0.005	S	0.002
Co	0.093	Sb	<0.05
Cr	20.3	Si	0.31
Cu	0.61	Sn	0.006
Fe	[53.1]	Ta	<0.005
Mg	<0.005	Ti	0.002
Mn	0.95	V	0.063
Mo	6.07	W	0.03
N	0.22	Zr	0.003
Nb	0.03		

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by April 20, 2022.