

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis BS 285BB

Certified Reference Material for Chill-cast Iron

	Estimated Analysis ¹		Estimated Analysis ¹
Analysis listed as percent by weight			
Al	0.01	Nb	0.004
As	0.002	Ni	1.4
B	0.008	P	0.05
C	3.4	Pb	<0.005
Ca	<0.005	S	0.01
Co	0.004	Si	1.9
Cr	1.0	Sn	0.002
Cu	0.3	Ti	0.04
Fe	90	V	0.1
Mg	0.046	W	0.06
Mn	0.7	Zr	0.006
Mo	0.2		

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 35 mm in diameter and 30 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by March 23, 2020.