

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis

### BS 360C

Certified Reference Material for Free Cutting Brass Grade CDA 360 - UNS Number S36000

	Estimated Analysis <sup>1</sup>		Estimated Analysis <sup>1</sup>
Analysis listed as percent by weight			
Ag	<0.005	Ni	0.019
Al	0.02	O	<0.005
As	0.004	P	0.011
Be	<0.005	Pb	2.8
C	0.001	S	<0.05
Co	<0.005	Sb	0.005
Cr	<0.005	Si	0.003
Cu	61.6	Sn	0.11
Fe	0.099	Zn	35.3
Mn	0.003		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

**A detailed final certificate of analysis will be supplied by November 21, 2021.**