

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis

BS 360D

Certified Reference Material for Free Cutting Brass Grade CDA 360 - UNS Number S36000

	Estimated Analysis ¹		Estimated Analysis ¹
Analysis listed as percent by weight			
Ag	<0.005	Ni	0.029
Al	<0.005	O	<0.005
As	0.009	P	0.002
Be	<0.005	Pb	2.7
C	0.001	S	<0.05
Co	<0.005	Sb	0.007
Cr	<0.005	Si	<0.005
Cu	61.8	Sn	0.15
Fe	0.16	Zn	35.1
Mn	0.001		

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by November 21, 2021.