

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis BS 400D

Certified Reference Material for Nickel Alloy Monel 400 - UNS Number N04400

Estimated Analysis <sup>1</sup>		Estimated Analysis <sup>1</sup>	
	Analysis listed as percent by weight		
Al	0.03	P	<0.005
B	<0.0005	Pb	<0.005
C	0.13	S	<0.005
Co	0.03	Sb	<0.005
Cr	0.002	Si	0.1
Cu	32.5	Sn	<0.005
Fe	2.0	Ta	<0.05
Mg	0.02	Ti	0.06
Mn	1.0	V	0.001
Mo	0.002	W	<0.005
Nb	<0.0005	Zr	0.0003
Ni	63.75		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

**A detailed final certificate of analysis will be supplied by June 30, 2014.**