

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis BS 430

Certified Reference Material for Stainless Steel Grade 430 - UNS Number S43000

	Estimated Analysis ¹		Estimated Analysis ¹
Analysis listed as percent by weight			
Al	<0.05	Nb	0.01
As	0.003	Ni	0.22
B	<0.005	O	(0.006)
C	0.061	P	0.024
Ca	<0.005	Pb	<0.005
Co	<0.5	S	0.002
Cr	16.5	Si	0.37
Cu	0.12	Sn	0.009
Fe	81.7	Ta	<0.05
Mg	<0.005	Ti	<0.05
Mn	0.71	V	0.057
Mo	0.052	W	<0.5
N	0.06	Zr	<0.005

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 44 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by September 19, 2018.