

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis

### BS 485A

Certified Reference Material for Naval Brass - UNS Number C48500

Estimated Analysis <sup>1</sup>		Estimated Analysis <sup>1</sup>	
Analysis listed as percent by weight			
Al	0.002	P	0.002
As	<0.005	Pb	1.4
C	0.001	S	<0.005
Cu	[61.7]	Sb	0.011
Fe	0.017	Si	0.001
Mn	0.002	Sn	0.73
Ni	0.002	Zn	36.1
O	<0.005		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

**A detailed final certificate of analysis will be supplied by December 5, 2021.**