

BS 61C analysis listed as percent concentration Certificate Number REV2-61C-012710

Analysis	C	Mn	P	S	Si	Cu	Ni	Cr	Mo	Al	As
1	0.181	0.743	0.012	0.0256	0.199	0.0279	0.536	0.495	0.167	0.029	0.0022
2	0.184	0.750	0.013	0.0256	0.200	0.028	0.55	0.498	0.167	0.030	0.0027
3	0.185	0.751	0.013	0.0258	0.200	0.0285	0.55	0.50	0.168	0.033	0.0028
4	0.188	0.753	0.015	0.0259	0.213	0.0292	0.55	0.509	0.17	0.034	0.0033
5	0.190	0.754	0.015	0.026	0.217	0.031	0.55	0.51	0.17	0.034	0.0036
6	0.190	0.766		0.028	0.217	0.031	0.552	0.51	0.17	0.035	
7	0.191	0.767			0.23	0.031	0.553	0.511	0.172	0.036	
8		0.775				0.034	0.555				
9							0.56				
Average	0.1870	0.7574	0.0136	0.0261	0.2109	0.0301	0.551	0.505	0.169	0.0330	0.00292
Std Dev	0.0037	0.0108	0.0013	0.0009	0.0117	0.0021	0.006	0.007	0.002	0.0026	0.00054
Certified	0.187	0.76	0.014	0.026	0.21	0.030	0.55	0.505	0.169	0.033	0.0029
t	2.45	2.36	2.78	2.57	2.45	2.36	2.31	2.45	2.45	2.45	2.78
C(95%)	0.0035	0.0090	0.0017	0.0010	0.0108	0.0017	0.005	0.006	0.002	0.0024	0.00068

continued from above

Analysis	Co	N	Sn	Sb	V	Ca	Nb	Pb
1	0.011	0.0047	0.0011	0.0002	0.0008	0.0004	<0.001	0.00002
2	0.011	0.0048	0.0013	0.00033	0.0010	0.0005	<0.002	0.00002
3	0.012	0.0050	0.0014	0.0004	0.0019		<0.002	0.00004
4	0.012	0.0053	0.0014	0.00041	<0.002			
5	0.0126		0.0018	0.00044	<0.002			
6	0.013			<0.002				
Average	0.0119	0.00495	0.00140	0.00036		0.00045		0.00003
Std Dev	0.0008	0.00026	0.00025	0.00010		0.00007		0.00001
Certified	0.012	0.0050	0.0014	0.0004	<0.002	(0.0004)	(<0.002)	(0.00003)
t	2.57	3.18	2.78	2.78				
C(95%)	0.0009	0.00042	0.00032	0.00012				

$C(95\%) = (t \times sd) / \sqrt{n}$ The half-width confidence interval, where t is the appropriate Student's t value, sd is the interlaboratory standard deviation, and n is the number of acceptable mean values. For further information regarding the confidence interval for the certified value see ISO Guide 35:2006 section 6.

Values in parentheses are not certified and are provided for information only.

Some of the co-operating laboratories were:

- Analytical Process Labs, Milwaukee, WI
- Brammer Standard Co., Inc., Houston, TX
- Chicago Spectro Service Laboratory, Chicago, IL
- Charles C. Kawin Company, Maywood, IL
- Crucible Specialty Metals, Syracuse, NY
- J. Dirats and Co., Inc., Westfield, MA
- LTV Steel, Aliquippa, PA
- Midstates Analytical Laboratories, Inc., Tulsa, OK
- VHG Laboratories, Inc., Manchester, NH

Analysis: Chemical analyses were made on millings from cross-sections of the bars. The individual values listed on the previous page are the average of each analyst's results.

Analytical Methods: Methods of analysis used were a combination of ASTM Standard Methods E 350, E 415, E 1019, plus additional ICP, and AA spectrometric methods.

Traceability: The following Certified Reference Materials were used to validate the analytical data listed above: NIST SRM 32e, 125b, 361 to 365; BAM 039-2, 044-1; BCS 455/1, 456/1, 458/1; ECRM 085-1, 088-1, 096-1, 184-1, 481-1; GBW 01402; IMZ 1.22, 1.74

Homogeneity: This Reference Material was tested for homogeneity using ASTM Standard Method E 826 and found acceptable. It was also examined by spark atomic emission spectrometry and found to be compatible with the following NIST Certified Reference Materials: SRM 1222, 1224, 1225, 1261A to 1265A, 1761 to 1767

Form: This Reference Material is machined in the form of a disc, approximately 41 mm diameter and 19 mm thick by Brammer Standard Company. The bar stock used for this material was produced by hot-rolling billets and annealing.

Certified area: The entire depth of the disc may be used.

Caution: As with any bar material, avoid spark atomic emission spectrometric burns in the center of the disc (5 mm radius), as some segregation may be present.

Sample Preparation: For best analytical results, use the same method for preparing the analytical surface on all reference materials as you use for production specimens. Avoid overheating the disc during surface preparation.

Certificate Number: The unique identification number for this certificate of analysis is REV2-61C-012710. This BS 61C Certificate of Analysis was revised to show the estimate of uncertainty for the certified values. After reviewing the analytical data, a third decimal place was certified for molybdenum and chromium plus a fourth decimal place was certified for arsenic. The niobium and lead values were changed to uncertified to conform to current versions of ISO Guides.

This Reference Material was first revised on March 20, 1992, after testing by additional laboratories in 1991 and 1992. The original Certificate of Analysis was dated May 30, 1987.

Refer to the "Certificates" section of the Brammer Standard Company website for any revision to this or other Brammer Standard Company's Certificates of Analysis.

Safety Notice: A Material Safety Data Sheet (MSDS) is not required for this material. This material will not release or otherwise result in exposure to a hazardous chemical, under normal conditions of use. Inquiries concerning this Reference Material should be directed to:

Brammer Standard Co., Inc. 14603 Benfer Road Houston, Texas 77069-2895 USA	Phone: (281) 440-9396 Fax: (281) 440-4432	website: brammerstandard.com email: contact@brammerstandard.com
--	--	--

Certified by: _____ on January 27, 2010.
Beau R. Brammer

Certificate Number REV2-61C-012710

Referenced Documents

ASTM documents available from ASTM, 1916 Race Street, Philadelphia, PA, 19103.

Versions available at the time of interlaboratory testing

E 350 - 90 Standard Test Methods for Chemical Analysis of Carbon Steel, Low-Alloy Steel, Silicon Electrical Steel, Ingot Iron, and Wrought Iron

E 415 - 85 (Reapproved 1989) Standard Test Method for Optical Emission Vacuum Spectrometric Analysis of Carbon and Low-Alloy Steel

E 826 - 85 (Reapproved 1990) Standard Practice for Testing Homogeneity of Materials for the Development of Reference Materials

E 1019-93 Standard Test Methods for Determination of Carbon, Sulfur, Nitrogen, and Oxygen in Steel and in Iron, Nickel, and Cobalt Alloys

ISO Guides and Standards available from Global Engineering - www.global.ihs.com

ISO Guide 35 Reference Materials - General and statistical principles for certification