

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis

BS 624

Certified Reference Material for Aluminum Bronze - UNS Number C62400

Estimated Analysis ¹		Estimated Analysis ¹	
	Analysis listed as percent by weight		
Al	10.2	P	0.003
As	0.005	Pb	0.002
C	0.004	S	0.001
Cu	[86.5]	Sb	0.004
Fe	3.02	Si	0.019
Mn	0.17	Sn	0.017
Ni	0.051	Zn	0.011
O	<0.005		

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 44 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by December 6, 2021.