

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis BS 6255

Certified Reference Material for Austenitic Nickel - UNS Number N06255

Estimated Analysis ¹		Estimated Analysis ¹	
	Analysis listed as percent by weight		
Al	0.09	Nb	0.10
B	0.002	Ni	[50.3]
C	0.012	O	<0.005
Ca	<0.05	P	0.013
Co	<0.5	S	0.0007
Cr	24.6	Si	0.27
Cu	0.76	Sn	0.001
Fe	16.5	Ta	<0.05
Mg	0.003	Ti	0.36
Mn	0.18	V	0.04
Mo	6.6	W	<0.5
N	<0.5	Zr	<0.05

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 63 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by September 20, 2021.