

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis BS 625E

Certified Reference Material for Inconel 625 - UNS Number N06625

Estimated  
Analysis<sup>1</sup>

Estimated  
Analysis<sup>1</sup>

Analysis listed as percent by weight

Al	0.21	Nb	3.6
As	<0.005	Ni	[60.6]
B	<0.05	O	<0.05
C	0.05	P	0.004
Ca	<0.005	Pb	<0.005
Co	0.03	S	<0.005
Cr	22.4	Si	0.07
Cu	<0.05	Ta	<0.5
Fe	3.8	Ti	0.26
Mg	<0.05	V	<0.05
Mn	0.06	W	<0.05
Mo	8.8	Zr	<0.05
N	<0.05		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

**A detailed final certificate of analysis will be supplied by October 26, 2018.**