

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis BS 690B

Certified Reference Material for Inconel 690 - UNS Number N06690

Estimated  
Analysis<sup>1</sup>

Estimated  
Analysis<sup>1</sup>

Analysis listed as percent by weight

Al	0.17	Nb	0.004
As	<0.005	Ni	[59.6]
B	<0.005	O	<0.005
C	0.03	P	0.003
Ca	<0.005	S	<0.005
Co	0.003	Si	0.21
Cr	29.6	Sn	<0.005
Cu	0.011	Ta	<0.05
Fe	10.1	Ti	0.21
Mg	0.004	V	0.01
Mn	0.11	W	0.001
Mo	<0.005	Zr	0.002
N	0.027		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 51 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by March 9, 2020.