

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis BS 70C

Certified Reference Material for Leaded Steel Grade 41L40 Mod

Estimated  
Analysis<sup>1</sup>

Estimated  
Analysis<sup>1</sup>

Analysis listed as percent by weight

Al	0.02	Nb	<0.05
As	<0.05	Ni	0.25
B	<0.005	O	<0.05
C	0.39	P	0.01
Co	0.009	Pb	0.13
Cr	0.97	S	0.02
Cu	0.12	Si	0.27
Fe	[96.5]	Sn	0.01
Mg	<0.005	Ti	0.002
Mn	0.90	V	0.002
Mo	0.20	Zr	<0.05
N	<0.05		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 41 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

**A detailed final certificate of analysis will be supplied by December 20, 2020.**

