

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis

BS 715B

Certified Reference Material for CDA 715 Copper Nickel - UNS Number C71500

Estimated Analysis ¹		Estimated Analysis ¹	
Analysis listed as percent by weight			
Al	0.006	O	<0.05
As	<0.005	P	0.003
Be	<0.005	Pb	0.004
C	0.015	S	0.005
Co	0.014	Sb	0.002
Cu	[67.3]	Si	0.011
Fe	0.59	Sn	<0.05
Mn	0.74	Zn	0.055
Ni	31.2	Zr	<0.005

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by November 5, 2021.