

Brammer Standard Company, Inc.

Certificate of Analysis

B.S. 750A

Nickel Alloy Type 750

Carbon	0.047	Molybdenum	0.22
Manganese	0.09	Vanadium	0.046
Phosphorus	(0.005)	Niobium	1.07
Sulfur	0.0007	Cobalt	0.29
Silicon	0.10	Titanium	2.60
Copper	0.04	Aluminum	0.74
Nickel	71.9	Iron	7.07
Chromium	15.68	Boron	0.0033

Some of the co-operating laboratories were:

Alpha Research Laboratories, Stevensville, MI
Allegheny Ludlum Steel Corp., Brackenridge, PA
Allegheny Ludlum Steel Corp., Lockport, NY
Brammer Standard Co., Houston, TX
Crucible Specialty Metals, Syracuse, NY
Jessop Steel Co., Washington, PA
Midstates Analytical Laboratories, Tulsa, OK

CAUTION: Because this Reference Material contains a high percent of chromium, iron, titanium, and niobium, care must be taken in its application. Corrections are likely to be necessary due to element interference.

See data on reverse side.

Certificate No. H750A-110689

Brammer Standard Company, Inc., 14603 Benfer Road, Houston, TX 77069
Telephone (281) 440-9396 Telex 775376 Fax (281) 440-4432