

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis BS 750D

Certified Reference Material for Inconel X750 - UNS Number N07750

Estimated
Analysis¹

Estimated
Analysis¹

Analysis listed as percent by weight

Al	0.70	Nb	0.89
As	<0.05	Ni	[71.3]
B	<0.05	O	<0.005
C	0.04	P	0.007
Co	0.03	S	<0.005
Cr	15.5	Si	0.19
Cu	0.03	Sn	<0.005
Fe	8.4	Ta	<0.5
Mg	<0.05	Ti	2.6
Mn	0.19	V	<0.5
Mo	<0.5	W	<0.0001
N	<0.05	Zr	<0.5

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by October 26, 2018.