

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis BS 825F

Certified Reference Material for Incoloy 825 - UNS Number N08825

Estimated  
Analysis<sup>1</sup>

Estimated  
Analysis<sup>1</sup>

Analysis listed as percent by weight

Al	0.1	Nb	0.02
B	0.002	Ni	[39.5]
C	0.01	O	<0.005
Ca	<0.005	P	0.01
Co	<0.5	S	0.0009
Cr	22.6	Si	<0.5
Cu	1.77	Sn	0.004
Fe	30.5	Ti	0.96
Mg	0.002	V	0.09
Mn	0.5	W	0.03
Mo	3.3	Zr	<0.005
N	<0.05		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

**A detailed final certificate of analysis will be supplied by September 19, 2018.**