

# BRAMMER STANDARD COMPANY, INC.

## Provisional Certificate of Analysis

### BS 836C

Certified Reference Material for Continuous Cast Leaded Red Brass Alloy Grade C836 - UNS Number C83600

Estimated Analysis <sup>1</sup>		Estimated Analysis <sup>1</sup>	
Analysis listed as percent by weight			
Ag	<0.5	Mn	<0.05
Al	0.002	Ni	0.37
As	0.008	O	<0.05
Be	<0.005	P	0.07
Bi	<0.5	Pb	4.8
C	0.005	S	0.037
Ca	<0.005	Sb	0.11
Co	<0.005	Si	0.003
Cu	[84.8]	Sn	4.8
Fe	0.02	Zn	4.9
Mg	<0.005	Zr	<0.005

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc approximately 44 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by October 14, 2021.