

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis

BS G30

Certified Reference Material for Hastelloy type G-30 - UNS Number N06030

Estimated Analysis ¹		Estimated Analysis ¹	
	Analysis listed as percent by weight		
Al	0.27	Ni	[45.4]
As	0.002	O	<0.05
B	0.001	P	0.01
C	0.005	Pb	<0.005
Ca	0.001	S	<0.005
Co	2.1	Sb	<0.005
Cr	28.9	Si	0.31
Cu	1.26	Sn	0.001
Fe	13.7	Ta	0.005
Mg	0.006	Ti	0.004
Mn	1.06	V	0.026
Mo	4.9	W	1.8
N	<0.5	Zr	0.002
Nb	0.42		

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by May 7, 2022.