

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis

BS H2E

Certified Reference Material for C-276 - UNS Number N10276

Estimated Analysis ¹		Estimated Analysis ¹	
	Analysis listed as percent by weight		
Al	0.4	N	0.013
As	<0.005	Nb	<0.05
B	<0.05	Ni	[58.3]
C	0.003	O	<0.005
Ca	<0.005	P	<0.03
Co	<0.5	S	<0.005
Cr	15.9	Si	<0.08
Cu	<0.05	Sn	<0.05
Fe	5.4	Ti	<0.05
Mg	<0.005	V	0.15
Mn	0.56	W	3.3
Mo	16.0	Zr	<0.005

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by July 07, 2019.