

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis BS LC-7A

Certified Reference Material for Commercially Pure Iron

	Estimated Analysis ¹		Estimated Analysis ¹
Analysis listed as percent by weight			
Al	0.019	Nb	0.0005
As	0.002	Ni	0.006
B	<0.0005	O	<0.05
C	0.002	P	0.014
Ca	0.0002	Pb	<0.005
Co	0.002	S	0.003
Cr	0.007	Sb	<0.005
Cu	0.004	Si	0.004
Fe	[99.9]	Sn	<0.005
Mg	<0.0005	Ti	0.0005
Mn	0.023	V	<0.005
Mo	0.003	W	<0.005
N	<0.05	Zr	<0.005

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 38 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by April 04, 2021