

Brammer Standard Company, Inc.
Certificate of Analysis

B.S. CE 24
AISI Carbon Steel Grade 1060 Reference Material

	Certified Value ¹	Estimate of Uncertainty ²		Certified Value ¹	Estimate of Uncertainty ²
Analysis listed as percent by weight					
C	0.63	0.01	Ni	0.045	0.002
Mn	0.84	0.02	Cr	0.016	0.002
P	0.016	0.002	Mo	0.006	0.0015
S	0.010	0.002	Al	0.044	0.003
Si	0.27	0.01	As	0.0029	0.0005
Cu	0.066	0.004	Co	0.010	0.002

¹ The certified value listed is the present best estimate of the true value based on the results of an interlaboratory testing program.

² The uncertainties listed are based on value judgments of the material inhomogeneity and possible bias in the determined analytical values. No attempt is made to derive exact statistical measurements of imprecision because several methods were used in the determination of most constituents.

The requirements of ISO Guide 31 and ISO Guide 35 were generally followed for the preparation of this reference material and certificate of analysis. This is a reference material as defined by ISO Guide 30.

See reverse side for more information.

Certificate Number CE24-030194

Brammer Standard Company, Inc., 14603 Benfer Road, Houston, TX 77069-2895
Telephone (281) 440-9396 Fax (281) 440-4432

Analysis	C	Mn	P	S	Si	Cu	Ni	Cr	Mo
1	0.614	0.82	0.015	0.0075	0.25	0.061	0.043	0.014	0.0043
2	0.618	0.83	0.015	0.008	0.26	0.064	0.044	0.015	0.0044
3	0.619	0.84	0.016	0.0088	0.26	0.0649	0.044	0.015	0.00533
4	0.62	0.84	0.016	0.009	0.26	0.065	0.044	0.016	0.0056
5	0.62	0.84	0.016	0.0091	0.26	0.065	0.0447	0.016	0.007
6	0.62	0.84	0.016	0.0094	0.263	0.0651	0.0448	0.016	0.0070
7	0.62	0.84	0.016	0.010	0.267	0.0656	0.045	0.016	
8	0.62	0.844	0.017	0.010	0.269	0.066	0.045	0.016	
9	0.621	0.844	0.017	0.010	0.27	0.066	0.045	0.016	
10	0.621	0.844	0.017	0.010	0.27	0.066	0.045	0.0160	
11	0.63	0.848	0.017	0.010	0.27	0.067	0.046	0.0163	
12	0.63	0.85	0.018	0.010	0.28	0.067	0.046	0.0165	
13	0.63	0.85	0.0184	0.011	0.28	0.068	0.046	0.017	
14	0.63	0.85		0.011	0.28	0.068	0.046	0.018	
15	0.64	0.86		0.011	0.28		0.047	0.018	
16	0.64	0.87		0.012	0.285		0.048	0.018	
17	0.648								
Average	0.626	0.844	0.0165	0.0098	0.269	0.0656	0.0452	0.0162	0.0056
Std.Dev.	0.009	0.011	0.0010	0.0012	0.010	0.0018	0.0012	0.0011	0.0012
Certified	0.63	0.84	0.016	0.010	0.27	0.066	0.045	0.016	0.006

continued from above

Analysis	Al	As	Co
1	0.041	0.0026	0.009
2	0.043	0.0027	0.0095
3	0.0435	0.0028	0.00956
4	0.0439	0.0029	0.010
5	0.046	0.0034	0.0104
6			0.012
Average	0.0435	0.0029	0.0101
Std.Dev.	0.0018	0.0003	0.0011
Certified	0.044	0.0029	0.010

Analysis: Chemical analyses were made on chips sampled from the bulk material. The laboratories participating in the testing normally followed the requirements of ISO Guide 25. The individual values listed above are the average of each analyst's results. Methods of analysis used were ASTM Standard Methods E 350 and E 1019 plus additional ICP and AA spectrometric methods.

Co-operating Laboratories: Some of the co-operating laboratories were:

Analytical Associates Inc., Detroit, Michigan
 Brammer Standard Co., Inc., Houston, Texas
 China Metallurgical Import & Export Corporation, Beijing, China
 Crucible Specialty Steel, Syracuse, New York
 J. Dirats and Co., Inc., Westfield, Massachusetts
 Shiva Technologies, Inc., Cicero, New York
 VHG Laboratories, Inc., Manchester, New Hampshire

Homogeneity: This Reference Material was tested for homogeneity using ASTM Standard Practice E 826 and found acceptable.

Traceability: The following Certified Reference Materials were used to validate the analytical data listed on page 2:
NIST SRM 5L, 16f, 32e, 121d, 122g, 125b, 361, 362, 363, 354, 365, 1270;
ECRM 085-1, 088-1, 096-1, 184-1, 481-1;
BCS 455/1, 456/1, 458/1; BAM 039-2, 044-1.

Source: This material was produced in by the China Metallurgical Import & Export Corporation, Beijing, China.

Available Form: This Reference Material is available only in the form of chips supplied in 150 g units.

Use: This Reference Material is intended for use in analytical chemistry. For information on uses of reference materials, refer to ISO Guide 33, NIST Special Publications 260-100 and 829.

Safety Notice: A Material Safety Data Sheet (MSDS) is not required for this material. This material will not release or otherwise result in exposure to a hazardous chemical under normal conditions of use. Inquiries concerning this Reference Material should be directed to:

Brammer Standard Co., Inc.
14603 Benfer Road
Houston, Texas 77069-2895 USA

Phone: (281) 440-9396

Fax: (281) 440-4432

Certified by: _____ on March 1, 1994.
G. R. Brammer

Certificate Number CE24-030194

References:

ASTM documents available from ASTM, 1916 Race Street, Philadelphia, PA, 19103.

E 350-90 Standard Test Methods for Chemical Analysis of Carbon Steel, Low-Alloy Steel, Silicon Electrical Steel, Ingot Iron, and Wrought Iron

E 826 - 85 (Reapproved 1990) Standard Practice for Testing Homogeneity of Materials for the Development of Reference Materials

E 1019-93 Standard Test Methods for Determination of Carbon, Sulfur, Nitrogen, and Oxygen in Steel and in Iron, Nickel, and Cobalt Alloys

ISO Guides available from American National Standards Institute, 11 West 42nd St., 13th Floor, New York, NY 10036.

ISO Guide 25 (Third edition, 1990), General requirements for the competence of calibration and testing laboratories.

ISO Guide 30 (Second edition, 1991), Terms and definitions used in connection with reference materials.

ISO Guide 31 (First edition, 1981), Contents of certificates of reference materials.

ISO Guide 33 (First edition, 1989), Uses of certified reference materials.

ISO Guide 35 (Second edition, 1989), Certification of reference materials - General and statistical principles.

Other useful documents available at no cost from NIST, U.S. Department of Commerce, Gaithersburg, MD 20899.

NBS Special Publication 260-100, Handbook for SRM Users

NIST Special Publication 829, Use of NIST Standard Reference Materials for Decisions on Performance of Analytical Chemical Methods and Laboratories

Certificate Number CE24-030194