

Brammer Standard Company, Inc.

Certificate of Analysis

BS HON U

Certified Reference Material for Carbon, Sulfur, Oxygen, Nitrogen, and Hydrogen in Steel

	Certified Value ¹	Estimate of Uncertainty ²	Certified Values³	Certified Value ¹	Estimate of Uncertainty ²
C	500	20			
H	1.0	0.5			
N	254	8			
O	23	5			
S	18	2			

For each element, the certified value listed is the present best estimate of the true value based on the mean of the weighted results of an interlaboratory testing program. See page 2 for more information on its calculation.

² For each element, the uncertainty listed is based on a statistical evaluation of the contributions of homogeneity and the interlaboratory testing program. See page 2 for more information on its calculation.

³ Values are given in parts per million.

The requirements of ISO Guides 30, 31, and 35 were followed for the preparation of this Certified Reference Material and certificate of analysis.

Analysis	*	C	*	H	*	N	*	O	*	S
1	1	429.67	2	0.5	2	244.8	2	11	1	14.33
2	1	480	2	0.775	2	246.7	2	17.33	1	15
3	1	481.33	2	0.83	2	247.67	2	18	1	15
4	1	483.33	2	0.8467	2	248.4	2	21	1	16.33
5	1	492.67	2	1.0	2	249.8	2	24.33	1	16.33
6	1	499	2	1.033	2	250.4	2	24.5	1	17
7	1	506.67	2	1.1	2	251	2	25.67	1	17
8	1	508.67	2	1.2	2	252	2	26	1	17.97
9	1	511	2	1.267	2	253.67	2	29.03	1	18.33
10	1	513.33	2	1.467	2	256.33	2	29.867	1	20
11	1	520	2	1.5	2	256.5	2	30.3	1	20.07
12	1	520	2	1.68	2	257	2	30.33	1	21
13	1	520	2	1.867	2	260	2	31.0	1	22
14	1	521.33	2	2.975	2	260	2	31.2	1	22
15	1	535	2	3.3	2	261	2	32.567	1	22.63
16	1	537.67	2		2	267.67	2	37.6		
Average		502.00		1.16		253.80		22.83		17.57
Std Dev		13.00		0.28		8.90		0.61		0.46
H		17.09		2.1700		12.209		4.6		4.220
U1		21.00		22.00		15.00		4.60		4.20
t-statistic		21300		21800		21314		21300		21400
U2		45.00		4.80		32.00		9.90		9.10
U3		11.00		1.30		8.00		2.50		2.40
Certified		500		1.0		254		23		18
Uncertainty		20		0.5		8		5		2
Tolerance		60		1.5		32		15		9

For each element, in accordance with the requirements of ISO Guides 34 and 35, an effort must be made to account for the effects on the certified value of the uncertainty estimate from homogeneity testing (H) and the uncertainties of the contributing laboratories. The average (A) is calculated using a weighted mean where the reciprocal of the square of each laboratory's combined uncertainty (C_L), calculated from its standard deviation (S_L) and its uncertainty estimate (U_L), is used as the weight (W_L) for its mean (M_L). The standard deviation (S) is calculated as the square root of the reciprocal of the sum of the weights. U_1 is the combined uncertainty from homogeneity and labs. U_2 is U_1 multiplied by the coverage factor (95 % t-statistic). U_3 is U_2 divided by the square root of the number of determinations (n). Thus:

$$C_L = \sqrt{S_L^2 + U_L^2} \quad W_L = \frac{1}{C_L^2} \quad A = \frac{\sum_{i=1}^n W_L M_L}{\sum_{i=1}^n W_L} \quad S = \frac{1}{\sqrt{\sum_{i=1}^n W_L}} \quad U_1 = \sqrt{H^2 + S^2} \quad U_2 = t \times U_1 \quad U_3 = \frac{U_2}{\sqrt{n}}$$

All but the final reported values are taken to two significant figures as determined by each quantity's uncertainty estimate. The final reported Uncertainty is U_3 rounded to one significant figure and represents the half width of the 95 % confidence interval for the **Certified** value. The final reported **Certified** value is A rounded to the same decimal place as the Uncertainty. The Uncertainty is a measure of the quality of the **Certified** value.

The Tolerance is a measure of the expected performance of an analysis. This involves further expanding the sample uncertainty to include instrument and operator uncertainty, for those without access to such calculations.

For further information regarding the confidence interval for the certified value see ISO Guide 35:2006 section 6.

Analytical Method Codes:

1 Combustion (ASTM E1019)	7 Photometric	13 Titrimetric
2 Fusion (ASTM E1019)	8 Flame Atomic Absorption	14 DCP Atomic Emission
3 Spark Atomic Emission	9 GF Atomic Absorption	15 HG Atomic Fluorescence
4 ICP Atomic Emission	10 X-Ray Fluorescence	16 Difference
5 ICP Mass Spectrometry	11 GD Atomic Emission	
6 Gravimetric	12 GD Mass Spectrometry	

ICP = Inductively Coupled Plasma GF = Graphite Furnace GD = Glow Discharge
 DCP = Direct Current Plasma HG = Hydride Generation

Lab Name	Location	Registrar	Accreditation
Brammer Standard Company, Inc.	Houston, TX	A2LA	17025, Guide 34
NSL Analytical	Cleveland, OH	ANAB	17025
Dirats Laboratories	Westfield, MA	ANAB	17025
LECO Corporation	St. Joseph, MI	A2LA	17025
Laboratory Testing, Inc.	Hatfield, PA	PRI/Nadcap	17025
Evans Analytical Group	Liverpool, NY	A2LA	17025
Exova	Glendale Heights, IL	A2LA	17025
Anderson Laboratories, Inc.	Greendale, WI	A2LA	17025
Instytut Metalurgii Zelaza	Gliwice, Poland	PCA	AB 554
TUV Rheinland India Pvt Ltd	Bangalore, India	NABL	17025
Elemental Analysis, Inc.	Lexington, KY	A2LA	17025
Luvak Inc.	Boylston, MA	PRI/Nadcap	17025
National Analysis Center For Iron And Steel	Beijing, China	CNAS	17025
IMR Test Labs, Surface Technologies	Lansing, NY	A2LA	17025

A2LA = American Association for Laboratory Accreditation
 ANAB = ANSI-ASQ National Accreditation Board
 CNAS = China National Accreditation Service
 NABL = National Accreditation Board for Testing and Calibration Laboratories
 Nadcap = National Aerospace and Defense Contractors Accreditation Program
 PCA = Polish Center For Accreditation
 PRI = Performance Review Institute

Analysis: Chemical analyses were made on 1g balls from representative samples for the certified portion of the lot in accordance with ASTM Standard Practice E1806. The laboratories participating in the testing followed the requirements of ISO Standard 17025.

Traceability: The following Certified Reference Materials were used to validate the analytical data: 13X31254; AR 115C, 546, 558, 612B, 644, 646, 649, 654, 657, 668, 670, 892, 961, 1650, 1652, 1653, 1656; BAS 451/1; BS CSN-1, CSN-4, HON-T, ST-626; JSS GS-1D; LECO 501-502, 501-529, 501-550, 501-674, 501-675, 501-676, 502-102, 502-197, 502-348, 502-416, 502-449, 502-868, 502-870, 502-874, 762-747, 802-869; SRM 15F, 36, 348A, 2159, 2166, 2171.

Homogeneity: This Certified Reference Material (CRM) was tested for homogeneity following ASTM Standard Method E826 and found acceptable. It was also examined by combustion/fusion and found to be compatible with the following Reference Materials — BAS 4/94; BS HON T; DSZU CA012; LECO 502-869, 502-870; NCS NS11043, NS20035B; Y 41340B.

Validity statement: ISO Guide 31 states that the certification should contain an expiration date for all materials where instability has been demonstrated or is considered possible, after which the certified value is no longer guaranteed by the certifying body. The certification of BS HON U is valid indefinitely. The certification is nullified if this CRM is damaged, contaminated, or otherwise modified.

Storage: This CRM must be stored in a cool, dry, non-corrosive environment.

Source: The stock for this CRM was produced by Tianjin D&C Technology Development; Tianjin, China.

Form: This CRM is in the form of a 1g balls.

Use: This CRM is intended for use in solution based methods of analysis. Refer to ISO Guide 33 for information about the use of Certified Reference Materials.

Certified Area: The entire CRM may be used.

Sample Preparation: For best analytical results, use the same method on all reference materials as used for production specimens.

Certificate Number: The unique identification number for this certificate of analysis is HON U-061317. You may obtain information on revisions of certificates from the internet at www.brammerstandard.com.

Safety Notice: A Safety Data Sheet (SDS) is not required for this material. This material will not release or otherwise result in exposure to a hazardous chemical, under normal conditions of use. Inquiries concerning this Reference Material should be directed to:

Brammer Standard Co., Inc.
14603 Benfer Road
Houston, Texas 77069-2895 USA

Phone: (281) 440-9396 Web: www.brammerstandard.com

Fax: (281) 440-4432 Email: contact@brammerstandard.com

Brammer Standard Company, Inc., is accredited by the American Association For Laboratory Accreditation (A2LA) to ISO Standard 17034 as a Reference Material Producer for the production of Certified Reference Materials and Reference Materials (Certificate Number 656.02)

Brammer Standard Company's Chemical Laboratory is accredited by A2LA to ISO Standard 17025. (Certificate Number 656.01)

By Certificate Number 10539, the Quality System of Brammer Standard Company, Inc., is registered to ISO 9001:2008 by National Quality Assurance (NQA), U.S.A.

The scopes of accreditation are listed on the website: www.brammerstandard.com

References:

Versions used were those available at the time of testing and characterization

E826 Standard Practice for Testing Homogeneity of a Metal Lot or Batch in Solid Form by Spark Atomic Emission Spectrometry

E1019 Standard Test Methods for Determination of Carbon, Sulfur, Nitrogen, and Oxygen in Steel, Iron, Nickel, and Cobalt Alloys by Various Combustion and Fusion Techniques

E1806 Standard Practice for Sampling Steel and Iron for Determination of Chemical Composition

ISO Standard 17025:2005 General requirements for the competence of testing and calibration laboratories

ISO Standard 9001:2008 Quality Management Systems - Requirements

ISO Guide 30:2015 Terms and definitions used in connection with reference materials + 2008 amendment

ISO Guide 31:2015 Reference materials - Contents of certificates and labels

ISO Guide 33:2015 Uses of certified reference materials

ISO Standard 17034:2016 General requirements for the competence of reference material producers

ISO Guide 35:2006 Reference Materials - General and statistical principles for certification

ASTM documents available from ASTM, 100 Barr Harbor Dr., West Conshohocken, Pa 19428.

ISO Guides and Standards available from Global Engineering - www.global.ihs.com

Other useful documents available from NIST, U.S. Department of Commerce, Gaithersburg, MD 20899.

NIST Special Publication 260-100, Handbook for SRM Users

NIST Special Publication 829, Use of NIST Standard Reference Materials for Decisions on Performance of Analytical Chemical Methods and Laboratories

Certified by: _____ on June 13, 2017.

Beau R. Brammer

President