

# Brammer Standard Company, Inc.

## Provisional Certificate of Analysis

### BS 1020A

Certified Reference Material for Carbon Steel Grade 1020 - UNS Number G10200

Analysis listed as percent by weight

	Estimated Analysis <sup>1</sup>		Estimated Analysis <sup>1</sup>
Al	0.002	Nb	<0.005
As	<0.05	Ni	0.060
B	0.0002	O	<0.05
C	0.22	P	0.014
Ca	0.001	Pb	<0.005
Co	0.006	S	0.023
Cr	0.12	Sb	<0.005
Cu	0.16	Si	0.23
Fe	[98.5]	Sn	0.017
H	<0.005	Ti	0.002
Mg	<0.005	V	0.035
Mn	0.58	W	0.005
Mo	0.018	Zr	<0.005
N	0.010		

<sup>1</sup> The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

**Form:** This CRM is machined in the form of a disc, approximately 38mm in diameter and 19mm thick by Brammer Standard Company, Inc.

**A detailed final certificate of analysis will be supplied by August 2, 2026**

1020A	Al	As	B	Be	Bi	C	Ca	Ce	Co	Cr	Cu	Fe	H	Mg
CSONH						0.224							0.00015	
BSC SAES	0.0019	0.0058	0.0002			0.228	0.0015		0.0062	0.115	0.161	98.49		0.0003
BSC GDS	0.0009		0.0001			0.227	0.0011		0.0053	0.116	0.163	98.51		
MTR	0.002		0.0003			0.21	0.0012			0.115	0.17			
Average	0.0016	0.0058	0.0002			0.22225	0.00127		0.00575	0.11533	0.16467	98.5	0.00015	0.0003
Certificate	0.002	<0.05	0.0002			0.22	0.001		0.006	0.12	0.16	[98.5]	<0.005	<0.005
1020A	Mn	Mo	N	Nb	Ni	O	P	Pb	S	Sb	Si	Sn	Ta	Ti
CSONH			0.0101			0.0074			0.0202					
BSC SAES	0.585	0.019			0.0595		0.0136		0.0223		0.225	0.0156		0.002
BSC GDS	0.583	0.0181		0.0004	0.0588		0.0132	0.0003	0.023	0.0017	0.223			0.0019
MTR	0.58	0.018	0.0098	0.002	0.062		0.014	0.0026	0.025		0.23	0.018		0.0008
Average	0.58267	0.01837	0.00995	0.0012	0.0601	0.0074	0.0136	0.00145	0.02263	0.0017	0.226	0.0168		0.00157
Certificate	0.58	0.018	0.010	<0.005	0.060	<0.05	0.014	<0.005	0.023	<0.005	0.23	0.017		0.002
1020A	V	W	Zn	Zr										
CSONH														
BSC SAES	0.0327	0.005												
BSC GDS	0.0345	0.0043		0.0007										
MTR	0.037													
Average	0.03473	0.00465		0.0007										
Certificate	0.035	0.005		<0.005										

**Homogeneity:** This Certified Reference Material (CRM) was tested for homogeneity using ASTM Standard Method E826 and found acceptable. It was also examined by spark atomic emission spectrometry and found to be compatible with the following Reference Materials: AR 1020; BS 57A, 57B, 57C, 57D, 57E, 1020, 1026A, 4951; JSS 412-1.

**Validity statement:** ISO Guide 31 states that the certification should contain an expiration date for all materials where instability has been demonstrated or is considered possible, after which the certified value is no longer guaranteed by the certifying body. The certification of BS 1020A is valid indefinitely. The certification is nullified if this CRM is damaged, contaminated, or otherwise modified.

**Storage:** This CRM must be stored in a cool, dry, non-corrosive environment.

**Source:** The bar stock for this CRM was produced by AMS Resource, Inc., McHenry, IL.

**Certified Area:** The entire depth of the CRM may be used.

Caution: As with any bar material, avoid spark atomic emission spectrometric burns in the center of the CRM (5 mm radius), as some segregation may be present.

**Sample Preparation:** For best analytical results, use the same method for preparing the analytical surface on all reference materials as used for production specimens. Avoid overheating the sample during surface preparation.

Caution: CRM contains significant insoluble soft metal inclusions. Surface smearing may occur. Spark atomic emission spectrometers may require extended preburns to compensate.

**Safety Notice:** A Safety Data Sheet (SDS) is not required for this material. This material will not release or otherwise result in exposure to a hazardous chemical, under normal conditions of use. Inquiries concerning this Reference Material should be directed to:

**Brammer Standard Co., Inc.**  
14603 Benfer Road  
Houston, Texas 77069-2895 USA

**Phone: (281) 440-9396**  
**Fax: (281) 440-4432**

**Web: [www.brammerstandard.com](http://www.brammerstandard.com)**  
**Email: [contact@brammerstandard.com](mailto:contact@brammerstandard.com)**

**Brammer Standard Company, Inc., is accredited by the American Association for Laboratory Accreditation (A2LA) to ISO Standard 17034 as a Reference Material Producer for the production of Certified Reference Materials and Reference Materials (our current Certificate Number 656.02 expires 01/31/2025)**

**Brammer Standard Company's Chemical Laboratory is accredited by A2LA to ISO Standard 17025. (Our current Certificate Number 656.01 expires 01/31/2025)**

**By current Certificate Number 10539 expiring 01/01/2027 the Quality System of Brammer Standard Company, Inc., is registered to ISO 9001 by National Quality Assurance (NQA), U.S.A.**

**The scopes of accreditation are listed on the website: [www.brammerstandard.com](http://www.brammerstandard.com)**

### **References:**

Versions used were those available at the time of testing and characterization

- E826 Standard Practice for Testing Homogeneity of a Metal Lot or Batch in Solid Form by Spark Atomic Emission Spectrometry
- E1019 Standard Test Methods for Determination of Carbon, Sulfur, Nitrogen, and Oxygen in Steel, Iron, Nickel, and Cobalt Alloys by Various Combustion and Fusion Techniques
- E1806 Standard Practice for Sampling Steel and Iron for Determination of Chemical Composition
  
- ISO Standard 17025:2017 General requirements for the competence of testing and calibration laboratories
- ISO Standard 9001:2015 Quality Management Systems - Requirements
- ISO Guide 30:2015 Terms and definitions used in connection with reference materials + 2008 amendment
- ISO Guide 31:2015 Reference materials - Contents of certificates and labels
- ISO Guide 33:2015 Uses of certified reference materials
- ISO Standard 17034:2016 General requirements for the competence of reference material producers
- ISO Guide 35:2017 Reference Materials - General and statistical principles for certification

*ASTM documents available from ASTM, 100 Barr Harbor Dr., West Conshohocken, PA 19428.*

*ISO Guides and Standards available from Global Engineering - [www.global.ihs.com](http://www.global.ihs.com)*

*Other useful documents available from NIST, U.S. Department of Commerce, Gaithersburg, MD 20899.*

NIST Special Publication 260-100, Handbook for SRM Users

NIST Special Publication 829, Use of NIST Standard Reference Materials for Decisions on Performance of Analytical Chemical Methods and Laboratories

Certified by: \_\_\_\_\_ on August 2, 2024.

Beau R. Brammer

President